

Für einen schnellen Start ...

IMMER ZUR HAND:
aus unkaputtbarer
Kunststoffolie zur
Lagerung im
Transportkoffer



Schritt 1

Legen Sie Ihrem Pferd ein festsitzendes Halfter an und setzen Sie ihm den Dichtring auf den Nasenrücken. Bringen Sie nun den Halfterriemen so an, dass der querführende Riemen über den Nasenrücken führt, und hängen Sie die Karabiner rechts und links am Halfter ein. Schieben Sie nun das Gehäuse auf den Dichtring auf und rasten Sie den Halfterriemen mit den Schnellverschlüssen am Gehäuse ein.



Schritt 2

Legen Sie eine Filtermembrane in den Filterkorb ein und bringen Sie diesen am Gehäuse an. Schieben Sie nun die Steuereinheit über die Führungsschienen, bis diese einrastet.



Schritt 3

Befüllen Sie nun die Mesh-Verneblereinheit mit der vorbereiteten Inhalationslösung und bringen Sie diese durch Verbinden des Kabels mit dem rechten oder linken Ausgang der Steuereinheit an. Schieben Sie nun die Abdeckhaube entlang der Führungsschienen auf, bis diese einrastet.



Schritt 4

Starten Sie die Steuereinheit durch ca. 2-sekündiges Drücken des Tasters über die seitliche Öffnung in der Abdeckhaube. Nach dem Inhalationsvorgang schaltet sich die Steuereinheit automatisch ab.



Schritt 5

Nach dem Inhalationsvorgang müssen die benutzten Komponenten vor dem Verstauen gereinigt und getrocknet werden.

Vorsicht

Die Kurzanleitung zum SaHoMa®-II Inhalationsset für Pferde ersetzt in keinem Fall die ausführliche Gebrauchsanweisung, deren Kenntnis und Beachtung vorausgesetzt wird. In der Gebrauchsanweisung finden Sie u. a. weitere Hinweise sowie Informationen zum Programmwechsel, dem Aufladen der SaHoMa®-II Steuereinheit und zur Reinigung der einzelnen Komponenten.

For a quick start ...

ALWAYS ON HAND:
Made of unbreakable
plastic film for storage
in carrying box



Step 1

Put a tight-fitting halter on your horse and place the sealing ring on the bridge of its nose. Now attach the halter strap so that the transverse strap passes over the bridge of the nose and hook the carabiners on the right and left of the halter. Now slide the housing on over the sealing ring and click the halter strap into place on the housing using the quick-release fasteners.



Step 2

Insert a filter membrane into the filter basket and attach it to the housing. Now slide the control unit over the guide rails until it clicks into place.



Step 3

Now fill the mesh nebulization unit with the prepared inhalation solution and attach it by connecting the cable to the right or left output of the control unit. Then slide the cover along the guide rails until it clicks into place.



Step 4

Start the control unit by pressing the button for approx. 2 seconds via the opening on the side of the protection cover. After the inhalation process, the control unit switches off automatically.



Step 5

In the end, the used components must be cleaned and dried before being stored.

Caution

The brief instructions for the SaHoMa®-II inhalation set for horses do not replace the detailed instructions for use in any case. Knowledge and observance of the instructions for use is assumed. In the instructions for use you will find, among other things, further instructions as well as information on changing programs, charging the SaHoMa®-II control unit and cleaning the individual components.